

Densah® Bur Tracking Log



	VPLTT	VPLTT-S	VT1525	VT1525-S	VT1828	VT1828-S	VS2228	VT2535	VT2535-S	VT2838	VT2838-S	VS3238	VT3545	VT3545-S	VT3848	VT3848-S	VS4248	VT4555	VT4858	VS5258
1																				
2																				
3																				
4																				
5																				
6																				
7																				
8																				
9																				
10																				
11																				
12																				
13																				
14																				
15																				
16																				
17																				
18																				
19																				
20																				

Densah[®] Burs Maintenance & Replacement

Surgical drills and burs should be replaced after 12-20 osteotomies* or sooner when they are dulled or worn.

Cutting and Densifying must be done under constant water irrigation. A pumping motion is required to prevent overheating. It is recommended to keep a spare set of Densah[®] Burs on hand in the event replacement is needed during a surgery.

*Chacon GE, Bower DL, Larson PR, et al. Heat production by three implant drill systems after repeated drilling and sterilization. J Oral Maxillofac Surg. 2006;64(2):265-269.